

# **CARLTON FORGE WORKS**

7743 E. ADAMS  
PARAMOUNT, CA 90723  
(562)633-1131

<b>CODE</b>  02	<b>SUBJECT</b>  <b>GAGE AND INSTRUMENT CALIBRATION</b>	<b>REVISION DATE:</b>  08/12/2014
<p><u>REVISIONS:</u></p> <p>09/09/97: ADDED ISO 10012-1</p> <p>08/25/03 – Re-written. Added Z540-1 and P&amp;W rqmts.</p> <p>12/19/05 – MIL-I-45662 changed to MIL-STD-45662</p> <p>12/08/11 – ANSI/NCSL Z540-1 changed to THE CURRENT REVISION OF ANSI/NCSL Z540</p> <p>08/12/14 – ISO 10012-1 (ANSI/NCSL Z540, MIL-STD45662... Changed to ISO 10012:2003, Removed MIL-STD45662</p>		

GAGE AND INSTRUMENT CALIBRATION SYSTEM MUST CONFORM TO THE CURRENT REVISION OF ISO 10012:2003 (ANSI/NCSL Z540, AS APPLICABLE). FOR PRATT & WHITNEY END USE, SYSTEM MUST CONFORM TO **THE CURRENT REVISION OF ANSI/NCSL Z540.**